

Ref. No.: GGSIPU/USCT/PROJECT/FRGS/PUR/2018/01

Date: 31.08.18

NOTICE INVITING QUOTATION

Sealed item rate quotations are invited from manufacturers/distributors/authorized dealers for the supply of following item with the following technical specifications, eligibility and terms and conditions:

Sr. No.	Items	Technical Specifications	Qty.
1.	Micrometer Adjustable Film Applicator	The specifications of " Micrometer Adjustable Film Applicator " are as follows: The film applicator is required for making thin film on glass plate. Film gap can be adjusted in 10 μ m steps by means of two integrated micrometric screws (0- 6 mm). <ol style="list-style-type: none"> 1. It should have 100 mm width. 2. It should make wet film thickness between 0 - 3500 μm with precision $\pm 2 \mu$m. 3. Applicator should be of made of stainless steel (SS 304) 4. Body of the block should be made of aluminum. 5. It should have two controllable micrometer. 6. It should have provision for cleaning after use. 	01

Eligibility:

1. Suppliers must have valid registration of GST
2. Should have valid PAN number.

Terms & Conditions:

Delivery: The item will be delivered to Dr. Biswajit Sarkar, Associate Professor, University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector: 16C, Dwarka, New Delhi-110078.

Delivery period: Within 30 days after receipt of PO. Supply after the stipulated delivery period will not be accepted.


Warranty: One year from the date of installation in good working conditions.

Taxes: Tax as applicable may please be clearly mentioned in the quotation.

Payment: After satisfactory Installation and demonstration. No advance payment will be made. Payment will be made through net banking.

Quotations received after due date & time shall be summarily rejected.

Last date, time and venue for submission of quotation	17.09.18, 5:00 PM, Room No. 105, Block B, USCT, GGSIPU, Sector 16 C, Dwarka, New Delhi - 110078
---	---


 Dr. Biswajit Sarkar
 Associate Professor, USCT