

UNIVERSITY SCHOOL OF BIOTECHNOLOGY
GURU GOBIND SINGH INDRAPRASTHA UNIVERSITY

Sec-16 C Dwarka, New Delhi.

Notice inviting Quotations

Suppliers are requested to quote the lowest rates for **Analytical Weighing Balance** with the following specifications:

Analytical weighing balance with maximum capacity of 220g and minimum readability of 0.01 mg. Accurate and precise with the internal automatic calibration and backlit graphic display. Minimum load that can be measured should be 10 mg and repeatability of 0.1 mg. Weighing area should be accessible with sliding door and sliding top glass. Operational power supply: 220-230 V. Additional features are desirable. Warranty: 2 years

Terms and Conditions:

1. Sales Tax/VAT as applicable may please be clearly mentioned.
2. No advance payment will be made. Payment will be released after satisfactory installation
3. The firm/s must be manufacturer or authorized supplier of the quoted equipment.

Your quote should reach the office of the undersigned latest by 5:00 pm on 8th December 2011.

Meenu Kapoor

Room No. AFR 106

University School of Biotechnology

Guru Gobind Singh Indraprastha University

Sec-16 C Dwarka, New Delhi